

OIML Member State
The Netherlands

Number R60/2000-A-NL1-19.03
Project number 2234345
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Hottinger Baldwin Messtechnik GmbH Im Tiefen See 45 D-64293 Darmstadt Germany
Identification of the certified type	A single point load cell , with strain gauges, Type : PW28...
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60 - Edition 2000 (E) for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
4 April 2019


C. Oosterman
Head Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2234345-01 dated 4 April 2019 that includes 51 pages;
- No. NMI-2234345-01 dated 4 April 2019 that includes 46 pages.

Characteristics of the load cell:

Maximum capacity (E_{max})	5 kg	10 kg up to and including 75 kg
Minimum dead load	0 kg	
Accuracy Class	C	
Rated Output	1,7 mV/V \pm 0,2 mV/V	2,0 mV/V \pm 0,2 mV/V
Maximum number of load cell intervals (n) ⁽¹⁾	4000	
Ratio of minimum LC Verification interval ⁽¹⁾ $Y = E_{max} / V_{min}$	25000	
Ratio of minimum dead load output return ⁽¹⁾ $Z = E_{max} / (2 * DR)$	9045	10572
Input impedance	400 Ω \pm 100 Ω	
Temperature range	-10 °C / + 40 °C	
Fraction p_{LC}	0,7	
Humidity Class	CH	
Safe overload	150 % of E_{max}	
Output impedance	400 Ω \pm 100 Ω	
Recommended excitation	5 V AC / DC	
Excitation maximum	15 V AC / DC	
Transducer material	Aluminum	
Atmospheric protection	IP67	

Remarks:

1. The characteristics for n_{max} , Y and Z can be reduced separately.

Each load cell produced is provided with an accompanying document with information about its characteristics.

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by the United States of America (NIST Handbook 44 and NCWM Publication 14), included in the Utilizer Declaration:

- R 60 OIML-CS rev.2 Additional requirements from the United States Accuracy class III L;
- R 60 OIML-CS rev.2 Additional requirements from the United States Marking requirements.