

Nederlands Meetinstituut

Member State  
The Netherlands

OIML Certificate N°R60/1991-NL-97.19  
Project number 10073806  
Page 1 of 2

### OIML CERTIFICATE OF CONFORMITY

**Issuing authority**

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1, Dordrecht  
Person responsible: A.J. Nederlof

**Applicant**

Name: Hottinger Baldwin Messtechnik GmbH  
Address: Im Tiefen See 45  
D-64293 Darmstadt  
Germany

**Manufacturer of the certified pattern**

Name: Hottinger Baldwin Messtechnik GmbH  
Address: Im Tiefen See 45  
D-64293 Darmstadt  
Germany

**Identification of the certified pattern**

Type : U2AD1  
 $P_i = 0.7$

Maximum Capacity ( $E_{max}$ )	200 kg , 500 kg, 1 t
Accuracy Class	D
Maximum number of load cell intervals (n)	1000
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	3500

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC5022 and the appertaining documentation folder, with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60  
edition 1991 (E)  
for accuracy class D

Nederlands Meetinstituut  
Hugo de Grootplein 1  
3314 EG Dordrecht (NL)  
Telephone +31 78 6332332  
Telefax +31 78 6332309


NMI B.V. (Registered at the Chamber of Commerce  
Delft number 26028701)  
Subsidiary companies:  
NMI Certin B.V. (26033418)  
NMI Van Swinden Laboratorium B.V. (26028703)  
NMI Inspecties en Kansspeltechniek B.V. (26028700)  
NMI International B.V. (26039176)

This certificate is issued under the provision that  
NMI B.V. nor its subsidiary companies accept any  
liability.

Reproduction of the complete certificate is allowed.  
Parts of the certificate may only be reproduced  
after written permission.



QUALIFIED  
BY STERLAB  
Reg. nr. L 029



Nederlands Meetinstituut

**Member State**  
The Netherlands

**OIML Certificate N°R60/1991-NL-97.19**  
Project number 10073806  
Page 2 of 2

This certificate relates only to the metrological and technical characteristics of the pattern of the Loadcell concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report number R60/1991-NL-97.19, that includes 38 pages.

The issuing authority  
A.J. Nederlof

*KW*

9 June 1997

The OIML member  
G.J. Faber

9 June 1997

\*  
\* \*

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.