

EC Test Certificate GB-1470 Revision 1

| Issued by: | National Measurement Office Notified Body Number 0126 | |
|--------------------------------|--|--|
| In accordance with | The Council Directive 2009/23/EC on Non-Automatic Weighing Instruments. Conformity with the essential requirements, referred to in Annex 1 of the Directive, is met by application of Paragraph 8.1 of the European standard on Metrological aspects of Non-Automatic Weighing Instruments EN45501:1992. | |
| | The applied error fraction p _i with reference to paragraph 3.5.4 of this standard is 0.5. | |
| Applicant: | Hottinger Baldwin Messtechnik GmbH Im Tiefen See 45 D-64293 Darmstadt GERMANY | |
| In respect of: | A model of an indicating device tested as a part of a weighing instrument. Manufacturer: See applicant Type: WE2111 digital indicator | |
| Characteristics: | Suitable for a Non-Automatic Weighing Instrument with the following characteristics: | |
| | n ≤ 10 000 for Class III instruments | |
| Description and documentation: | The Indicating device is described in the Descriptive Annex. Documents appertaining to this EC Test Certificate are held by | |

documentation: Documents appertaining to this EC Test Certificate are held by the National Measurement Office.

Remarks: The Indicator has been tested and found to conform to the relevant parts of EN45501 and WELMEC Guide 2.1. A summary of the tests performed in support of this EC Test Certificate is provided in the Appendix to the Descriptive Annex.

This revision replaces previous versions of the certificate.

 Date:
 04 April 2014

 Reference №:
 TS1201/0040

Signatory: P R Dixon for Chief Executive

National Measurement Office | Stanton Avenue | Teddington | TW11 0JZ | United Kingdom Tel +44 (0)20 8943 7272 | Fax +44 (0)20 8943 7270 | Web www.bis.gov.uk/nmo The NMO is an Executive Agency of the Department for Business, Innovation and Skills



National Measurement Office

TEST CERTIFICATION № GB-1470

Descriptive Annex

1 INTRODUCTION

This indicating device is designated the WE2111 digital indicator. It is designed to be used as part of a single/dual range/interval, Class III Non-Automatic Weighing Instrument. The indicator is self-indicating and DC or mains-powered.

2 DESCRIPTION

2.1 Construction

- **2.1.1** The Operator Panel (Figure 1) has the following features:
 - ABS plastic enclosure
 - Six digit LED display
 - Five functions keys
 - Weighing status
 - Multiple range/interval status
 - Connections and ports located at the rear

2.2 Devices

- Extended indicating device
- Printing device
- Totalising device
- Initial zero setting device ($\leq 20\%$ of Max)
- Zero tracking device ($\leq 4\%$ of Max)
- Semi-automatic zero setting device (≤ 4% of Max)
- Tare setting device: Semi-automatic, additive and subtractive
- Preset tare device
- Multi-range device
- Multi-interval device
- lb/kg switching device
- Gross/net switching device
- Calibration device
- Piece counting device
- Fault handling device
- Display test device, on power up
- Alibi memory device
- Zero indicator
- Indication of stable equilibrium

2.3 Operation

2.3.1 Switch-on

At switch on the operator panel will perform internal operation checks while displaying the initiation screen. This screen will show the software version number, the unique calibration number provided for the audit trail and the electronics serial number. On completion of the initialisation, the operator panel will enter the normal operating mode, and perform an initial zero-setting operation. The initial zero-setting range is \leq 20% of maximum capacity.

2.3.2 Zero-tracking

Zero tracking operates provided that the instrument is within range of not more than 4% of its capacity, and that the weight display is stable. The rate of correction is set not to exceed 0.5 d/s.

2.3.3 Semi-automatic zero setting

The zero key operates provided that the instrument is within range of not more than 4% of its capacity, and that the weight display is stable. Annuniciators are provided to indicate when the instrument is stable and at zero.

2.3.4 Over-range and under-range

If the load is less than gross zero, then the display shows U-----.

The instrument is set to display weights up to Max. At greater loads the display shows O----.

2.3.5 Additional functions

The keys located below the operator display have the following functions:

The WE2111 digital indicator has 5 front panel keys that control the operation of the instrument. The 6th key (SETUP) is on the rear of the instrument. The setup key can be sealed to prevent unauthorized tampering of trade critical settings and calibration.

Each of the front panel keys has two separate functions:

a normal function that is available during normal weighing (as printed on the key).

a setup function which is available during setup and calibration (as printed beneath the key).

3 TECHNICAL DATA

3.1 The WE2111 digital indicator has the following technical characteristics:

| Maximum number of scale intervals | 3 000 | 6 000 | 10 000 |
|---|---------------|-------|--------|
| Load cell excitation voltage | 5 V DC | | |
| Minimum load cell impedance | 21 Ω | | |
| Maximum load cell impedance | 5000 Ω | | |
| Minimum input voltage per verification scale interval | 0.5 µV/Div | | |
| Measuring range minimum voltage | 0.001 mV | | |
| Measuring range maximum voltage | 25 mV | | |
| Fraction of maximum permissible error | 0.5 | | |
| Operating temperature range | -10 / + 40 °C | | |
| Load cell connection | 6 wire | | |
| Load cell cable length m/mm ² (junction box to indicator) 755 378 22 | | 227 | |

Note: Cable length obtained from manufacture.

3.2 Documentation/Drawings

| Drawing reference № | Description |
|---------------------|----------------------|
| WE21-400-151 | Bill of materials |
| WE21-103-141 | Circuit Schematics |
| WE21-104-143 | |
| WE21-111-140 | |
| WE21-003-140 | Circuit Layout |
| WE21-004-142 | |
| WE21-011-140 | |
| WE21-B01-110 | Hardware description |
| WE21-B02-100 | Software Description |
| WE21-600-111 | Manual |
| WE21-A03-100 | Mechanical assembly |
| WE21-B05-100 | Sealing Description |
| R400-534-101 | Alibi Firmware |

Note: Model WE2111 is compatible with the C510 model as detailed in the above documentation/drawings.

3.3 Power supply

The indicator can either operate directly from mains AC supply (86-260 V AC 48-62 Hz) or via a stable DC supply (12-24 V DC). Any compatible CE-marked mains adaptor may be used.

4 PERIPHERAL DEVICES

The following peripheral devices may be connected to the interfaces provided:

- Peripheral devices that have been issued with a EC Test Certificate by a Notified Body responsible for type approval under Directive 2009/23/EC; or
- Peripheral devices without a EC Test Certificate under the following conditions:
 - it bears the CE marking for conformity to the EMC Directive;
 - it is not capable of transmitting any data or instruction into the weighing instrument, other than to release a printout, checking for correct data transmission or validation;
 - it prints weighing results and other data as received from the weighing instrument without any modification or further processing;
 - it complies with the applicable requirements of EN45501, i.e. 4.2, 4.4, 4.6 and 4.7.

A printing device may print additional information such as date or number to identify the printed weighing result(s) or sets of weighing results.

5 SOFTWARE

5.1.1 Change log

The WE2111 digital indicator contains a change log that records trade significant events. These include changes to trade relevant settings, creation, and clearing of the change log and trade relevant firmware upgrades. Only trade relevant settings are stored in the change log.

The change log records the following information about each change:

- The calibration counter at the time of the change
- The date and time of the change
- The system database name & setting name that has changed
- The menu name of the setting that has changed
- The index of the setting that has changed. This is only used for array settings, such as resolution where there is a resolution for each range.
- such as resolution where there is a resolution for each range
- Previous value of the setting New value of the setting
- The log mask

5.1.2 Change log Security

The change log is stored on the internal WE2111 filesystem which is not accessible, and hence cannot be tampered with. The WE2111 application is the only means by which to access the change log. The WE2111 contains Alibi software which allows the change log to be viewed. The Alibi change log viewer is part of the trade approval.

When a USB disk is attached to the WE2111, a copy of the change log can be made to the USB disk. This copy of the change log can always be verified against the internal change log, via the Alibi change log viewer.

5.1.3 Digital storage device (DSD)

The WE2111 contains a digital storage device (DSD) to record traceable readings. The DSD records the following information about each entry:

- \rightarrow The print ID
- \rightarrow The weight reading
- \rightarrow The tare weight
- \rightarrow The date and time

5.1.4 Format and capacity

The DSD is stored in a binary format internally. The DSD has a maximum size of 512 kilobytes, and can contain approximately 11900 records. The WE2111 can be configured to either prompt the user to purge 10% of records from the DSD, or automatically purge 10% of records, once the DSD is full. Records can also be manually purged from the DSD from the DSD:PURGE setup menu.

5.1.5 DSD Change log security

The DSD is stored on the internal WE2111 filesystem which is not accessible, and hence cannot be tampered with. The WE2111 application is the only means by which to access the DSD. The WE2111 contains Alibi software which allows the DSD records to be viewed. The Alibi DSD viewer is part of the trade approval.

When a USB disk is attached to the WE2111, the DSD can be exported to the USB disk in a comma separated value (CSV) file. CSV files can be easily viewed in most spreadsheet programs. This copy of the DSD can always be verified against internal DSD, via the Alibi DSD viewer.

6 VERIFICATION INFORMATION

6.1 Software identification

The legally relevant software (alibi) is designated v1.0x (where x refers to the identification of the non-legally relevant part of the software, which may be modified by the manufacturer).

The application software (non-legally relevant) is designated Pxxx or Vx.xx (where xxx and x.xx may be modified by the manufacturer).

7 MARKINGS

The instruments shall bear the following legends:

Manufacturer's mark or name Serial number EC Test Certificate number

8 ALTERNATIVES

There are no authorised alternatives.

9 ILLUSTRATIONS

Figure 1 WE2111 indicator

TEST CERTIFICATE HISTORY

| Issue №. | Date | Description |
|--------------------|---------------|-------------------------------------|
| GB-1470 | 15 July 2013 | Test Certificate first issued. |
| GB-1470 Revision 1 | 04 April 2014 | Test Report SN 1278 added to Annex. |

APPENDIX TO DESCRIPTIVE ANNEX

TESTS CARRIED OUT

The following tests were performed with the indicator connected to a load cell simulator or to a weighing platform.

| EN45501 Ref | Test | Report number |
|-------------|---|---------------|
| A.4.6.1 | Tare | NMO TR 627 |
| A.4.10 | Repeatability | NMO TR 627 |
| A.5.2 | Warm-up | NMO TR 627 |
| A.5.3.1 | Weighing performance at static temperatures | NMO TR 627 |
| A.5.3.2 | Temperature effect on no load indication | NMO TR 627 |
| B.2.2 | Damp heat steady state | NMO TR 627 |
| B.4 | Span stability | NMO TR 627 |

| OIML R76 (2006) | Test | Report number |
|-----------------|--|---------------|
| A.5.4 | Voltage variations | NMO SN 1236 |
| B.2.2 | Damp heat steady state | NMO SN 1236 |
| B.3.1 | Short time power reductions | NMO SN 1236 |
| B.3.2 | Bursts | NMO SN 1236 |
| B.3.3 | Electrostatic discharges | NMO SN 1236 |
| B.3.5 | Immunity to radiated electromagnetic fields | NMO SN 1236 |
| B.3.6 | Immunity to conducted radio-frequency fields | NMO SN 1236 |
| - | Resistance cable length test | NMO SN 1236 |

| OIML R76 (2006) | Test | Report number |
|-----------------|---|---------------|
| B.3.3 | Surges | NMO SN 1278 |
| B.3.1 | Short time power reductions | NMO SN 1278 |
| B.3.2 | Bursts | NMO SN 1278 |
| B.3.4 | Electrostatic discharges | NMO SN 1278 |
| B.3.5 | Immunity to radiated electromagnetic fields | NMO SN 1278 |
| B.3.6 | Immunity to conducted radio-frequency fields | NMO SN 1278 |
| B.3.7 | Electrical transients on instruments powered from a road vehicle power supply | NMO SN 1278 |

Note:

- EMC testing conducted at the levels required in OIML R76 (2006)
- Model WE2111 is compatible with the C510 model as detailed in the above test reports.



