

Nederlands Meetinstituut

**Member State**  
The Netherlands

**OIML Certificate N°R60/1991-NL-95.18**

Project number 10049409

Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI IJkwezen B.V.  
Address: Hugo de Grootplein 1, Dordrecht  
Person responsible: M. Charité

### Applicant

Name: HBM Inc.  
Address: 19 Bartlett Street  
Marlboro, Massachusetts 01752  
United States of America

### Manufacturer of the certified pattern

Name: HBM Inc.  
Address: 19 Bartlett Street  
Marlboro, Massachusetts 01752  
United States of America

### Identification of the certified pattern

Type: RSC  
 $P_i = 0.7$

Maximum Capacity ( $E_{max}$ )	100, 200, 300, 500, 750, 1000, 2000 and 5000 lb or 50, 100, 200, 500, 1000, 2000 and 5000 kg or 500, 1000, 2000, 5000, 10000, 20000 and 50000 N
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Minimum load cell verification interval ( $V_{min}$ )	$E_{max} / 8333$

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC2631 Revision 0 and the appertaining documentation folder, with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R60  
edition 1991 (E)  
for accuracy class C and D

**Nederlands Meetinstituut**  
Hugo de Grootplein 1  
3314 EG Dordrecht (NL)  
Telephone +31 78 6332332  
Telefax +31 78 6332309

Nederlands Meetinstituut N.V. (Registered at the Chamber of Commerce Delft number 28701)


Subsidiary companies:  
NMI Certin B.V. (33418)  
NMI Van Swinden Laboratorium B.V. (28703)  
NMI IJkwezen B.V. (28700)

This certificate is issued under the provision that Nederlands Meetinstituut N.V. nor its subsidiary companies accept any liability.

Reproduction of the complete certificate is allowed. Parts of the certificate may only be reproduced after written permission.



**QUALIFIED  
BY STERLAB**  
Reg. nr. L 029



Nederlands Meetinstituut

**Member State**  
The Netherlands

**OIML Certificate N°R60/1991-NL-95.18**  
Project number 10049409  
Page 2 of 2

This certificate relates only to the metrological and technical characteristics of the pattern of the Loadcell concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports:  
N° R60/1991-NL-95.18A, that includes 37 pages;  
N° R60/1991-NL-95.18B, that includes 37 pages.

The issuing authority  
M. Charité

6 December 1995

The OIML member  
G.J. Faber

6 December 1995

\*  
\* \*

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.